

CERTIFICATE OF STERILITY

Quality Certificate



Gamma Irradiation for CleanMax® Clean Sterile Cleanroom Apparel

| Sterility Lot Number | Product Release Date: | | Sterility Cert No. | |
|----------------------|-----------------------------------|---------------|--------------------|------------|
| CS20250919-US | 7th December 2025 | | 2502857-1 | |
| Dose Info | Min Specified | Max Specified | Min Actual | Max Actual |
| (KgY) | 27.5 | 42 | 27.5 | 40.3 |
| Process Code | M2502101/2102/2103/2104/2105/2106 | | | |
| ITEM NUMBER(S) | SIZE(S) | | | |
| CTL414CS | SM, MD, LG, XL | | | |
| CTL417CS | MD, LG, XL | | | |

Lakeland Industries Inc, doing business as "Lakeland Fire and Safety", certifies that the product/item number(s) and lot number(s) identified above have been processed and released in accordance with all relevant specifications and standard operating procedures.

The minimum validation dose has been established in accordance with guidelines outlined in ANSI/AAMI/ISO 11137-1 Sterilization of Health Care Products – Requirements for Validation and Routine Control – to provide a Sterility Assurance level of 10^{-6} . Do not use if package integrity has been compromised.

Product has a shelf-life of 5 years and will remain sterile for that period from the production date as long as the integrity of the packaging remains uncompromised.

Certificates of Sterility are available online www.lakeland.com or contact Customer Services at info@lakeland.com

D. Rumble, BSc, PCQI

Global Compliance and Product Manager

